

Kitemark™ Certificate

This is to certify that:

Interster International B.V.
Samsonweg 2
Wormerveer
1521 RC
Netherlands

Holds Certificate Number:

KM 633144


In respect of:

ISO 11140-1
Sterilization of Health care products. Chemical indicators

This issues the right and licence to use the Kitemark in accordance with the Kitemark Terms and Conditions governing the use of the Kitemark, as may be updated from time to time by BSI Assurance UK Ltd (the "Conditions"). All defined terms in this Certificate shall have the same meaning as in the Conditions.

The use of the Kitemark is authorized in respect of the Product(s) detailed on this Certificate provided at or from the above address.

For and on behalf of BSI:



Gary Fenton, Global Product Certification Director

First Issued: 14/09/2015

Latest Issue: 19/11/2015



Page: 1 of 3

...making excellence a habit.™

Kitemark™ Certificate

No. KM 633144

BS EN ISO 11140-1:2014 - Sterilization of health care products. Chemical indicators.

Model:	ISP® Steam Emulating Indicator
Product codes:	3FSKS645834 3FSKS645836
Description:	Steam Emulating Indicator. Sterilisation conditions change the ink colour from blue to light pink. Pack size - length: 25mm; width: 50mm.
Classification:	Sterilization chemical indicator Type 6.
Technical Specification:	BS EN ISO 11140-1:2014 Sterilization of health care products. Chemical indicators. General requirements.

Ink Timing

Product Code	Temperature	Stated value in decimal minutes
3FSKS645834	134°C	3.5 min
3FSKS645836	134°C	5.3 min

First Issued: 14/09/2015

Latest Issue: 19/11/2015

Page: 2 of 3

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 845 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

Kitemark™ Certificate

No. KM 633144

BS EN ISO 11140-4:2007 - Sterilization of health care products. Chemical indicators. Class 2 indicators as an alternative to the Bowie and Dick-type test for detection of steam penetration.

Model:	ISP® Bowie and Dick Type Indicator	
Product code:	3FSKS610860	
Description:	Bowie Dick type single use test pack.	
Classification:	Sterilization indicator Class 2	
Technical Specification:	BS EN ISO 11140-4:2007 Sterilization of health care products. Chemical indicators.	
Maximum Exposure Period: (decimal minutes)	Temp 134°C	3 - 3.5 min

First Issued: 14/09/2015

Latest Issue: 19/11/2015

Page: 3 of 3

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 845 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.